



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

**Scales Sales & Service, Inc.
8615 Whitmore Circle Suite 104
Omaha, Nebraska 68122**

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

*This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(As outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):*

**Calibration of Weighing Devices at Field Sites Controlled by the Laboratory
(As detailed by the supplement)**

Such testing and/or calibration services shall only be offered at or from the address given above. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

The validity of this certificate is mandated through ongoing surveillance.

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:
March 05, 2005

Accreditation No.:
59316

Issue Date:
January 14, 2011

Certificate No.:
L11-9

Expiration Date:
January 13, 2013

Page No.:
Page 1 of 2



Certificate of Accreditation: Supplement

Scales Sales & Service, Inc.
8615 Whitmore Circle Suite 104
Omaha, Nebraska 68122

Accreditation is granted to this facility to perform the following calibrations:

Mass, Force, and Weighing Devices

CALIBRATION FIELD	MEASURED QUANTITY, INSTRUMENT OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)
Analytical Balance	1 mg to 199 g	$(1.22 \times 10^{-4} \times 3.00 \times 10^{-6} \text{ wt}) \text{ g}$	Class 1
	200 g to 799 g	$(1.14 \times 10^{-2} \times 1.00 \times 10^{-6} \text{ wt}) \text{ g}$	
	800 g to 5 999 g	$(1.15 \times 10^{-1} \times 1.00 \times 10^{-6} \text{ wt}) \text{ g}$	
	6 000 g to 9 999 g	$(1.16 + 1.00 \times 10^{-6} \text{ wt}) \text{ kg}$	
Bench Scales	1 lb to 2 500 lb	$(5.78 + 3.00 \times 10^{-6} \text{ wt}) \text{ lb}$	Class F
Crane Scales	1 lb to 50 000 lb	$(11.55 + 2.80 \times 10^{-5} \text{ wt}) \text{ lb}$	
Counting Scales	1 lb to 25 000 lb	$(5.78 + 2.80 \times 10^{-5} \text{ wt}) \text{ lb}$	
Floor Scales	1 lb to 50 000 lb	$(11.55 + 2.80 \times 10^{-5} \text{ wt}) \text{ lb}$	
Hopper Scales	5 lb to 250 000 lb	$(23.1 + 5.60 \times 10^{-5} \text{ wt}) \text{ lb}$	
Truck Scales	250 lb to 240 000 lb	$(23.09 + 5.50 \times 10^{-5} \text{ wt}) \text{ lb}$	
Railroad Track Scales	2 500 lb to 400 000 lb	$(57.64 + 4.10 \times 10^{-5} \text{ wt}) \text{ lb}$	

1. Remarks: This column shall include pertinent information about the calibration of the Measured Instrument or parameter. The information should include the type of standards used and any pertinent information about the measurement method. This column is not to be used for commercial advertisement of laboratory services.
2. The term wt represents weight in pounds or grams (including SI multiple and submultiples units) appropriate to the uncertainty statement.